



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/4849227
Dated/दिनांक : 09-04-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	30-04-2024 09:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	30-04-2024 09:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Pmo
Department Name/विभाग का नाम	Department Of Atomic Energy
Organisation Name/संगठन का नाम	Uranium Corporation Of India Limited
Office Name/कार्यालय का नाम	Jaduguda
Total Quantity/कुल मात्रा	45
Item Category/मद केटेगरी	Power Contactors as per IS 13947 - 4 - 1 (Q3)
Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष	1 Year (s)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance/विगत प्रदर्शन	30 %
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Primary product category	Power Contactors as per IS 13947 - 4 - 1
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days

Bid Details/बिड विवरण**Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)**

No

Evaluation Method/मूल्यांकन पद्धति

Total value wise evaluation

EMD Detail/ईएमडी विवरण

Required/आवश्यकता

No

ePBG Detail/ईपीबीजी विवरण

Required/आवश्यकता

No

Splitting/विभाजन

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता

No

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of

primary product having highest value should meet this criterion.

4. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for 25%(selected by Buyer) percentage of total QUANTITY. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

5. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Power Contactors As Per IS 13947 - 4 - 1 (5 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	Length	162.0 (millimeter)
	Width	79.0 (millimeter)
	Height	120.0 (millimeter)
Generic	Auxiliary Contact	1 NO+1 NC
	Mounting	Surface Mounting Via Screw
	AC Rating (Ampere)	110
	Coil Voltage	415
	Contact Voltage	415VAC
	Add on Block	Side
	Duty	AC-3
Warranty Details	Time Duration for Repairing /Replace the defect during Warranty	With Two Week
	Warranty	6.0 (month)
Standards and Certifications	Standard compliance	IEC 60947-4-1, IS 13947-4-1
	Certification	ISO
Test Reports	Availability Of Test Report From Central Govt/Nabl/Ilac Accredited Lab To Prove Conformity To Specification	Yes

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	NA

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Supriyo Sarkar	832102,UCIL JADUGUDA STORE CHIEF SUPERINTENDENT STORE AT/PO-JADUGUDA MINES MAIN BUILDING URANIUM CORPORATION OF INDIA LTD Ph No : 9431139916	5	90

Power Contactors As Per IS 13947 - 4 - 1 (10 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	Length	180.0 (millimeter)
	Width	100.0 (millimeter)
	Height	60.0 (millimeter)
Generic	Auxiliary Contact	1 NO+1 NC
	Mounting	Surface Mounting Via Screw
	AC Rating (Ampere)	63
	Coil Voltage	415
	Contact Voltage	480V AC, 50-60 Hz
	Add on Block	Front
	Duty	AC-3

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Warranty Details	Time Duration for Repairing /Replace the defect during Warranty	With Two Week
	Warranty	12.0 (month)
Standards and Certifications	Standard compliance	IEC 60947-4-1, IS 13947-4-1
	Certification	NA
Test Reports	Availability Of Test Report From Central Govt/Nabl/Ilac Accredited Lab To Prove Conformity To Specification	No
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	NA

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Supriyo Sarkar	832102,UCIL JADUGUDA STORE CHIEF SUPERINTENDENT STORE AT/PO-JADUGUDA MINES MAIN BUILDING URANIUM CORPORATION OF INDIA LTD Ph No : 9431139916	10	90

Power Contactors As Per IS 13947 - 4 - 1 (10 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	Length	70.0 (millimeter)
	Width	40.0 (millimeter)
	Height	30.0 (millimeter)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Generic	Auxiliary Contact	1 NO+1 NC
	Mounting	Din Rail
	AC Rating (Ampere)	32
	Coil Voltage	415
	Contact Voltage	480V AC, 50-60 Hz
	Add on Block	Front
	Duty	AC-3
Warranty Details	Time Duration for Repairing /Replace the defect during Warranty	With Two Week
	Warranty	6.0 (month)
Standards and Certifications	Standard compliance	IEC 60947-4-1
	Certification	ISO
Test Reports	Availability Of Test Report From Central Govt/Nabl/Ilac Accredited Lab To Prove Conformity To Specification	Yes
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Supriyo Sarkar	832102,UCIL JADUGUDA STORE CHIEF SUPERINTENDENT STORE AT/PO-JADUGUDA MINES MAIN BUILDING URANIUM CORPORATION OF INDIA LTD Ph No : 9431139916	10	90

Power Contactors As Per IS 13947 - 4 - 1 (10 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	Length	83.0 (millimeter)
	Width	45.0 (millimeter)
	Height	88.0 (millimeter)
Generic	Auxiliary Contact	1NO
	Mounting	Din Rail
	AC Rating (Ampere)	22
	Coil Voltage	415
	Contact Voltage	415VAC
	Add on Block	Front
	Duty	AC-3
Warranty Details	Time Duration for Repairing /Replace the defect during Warranty	NA
	Warranty	18.0 (month)
Standards and Certifications	Standard compliance	IEC 60947-4-1, IS 13947-4-1
	Certification	NA
Test Reports	Availability Of Test Report From Central Govt/Nabl/Ilac Accredited Lab To Prove Conformity To Specification	Yes
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Supriyo Sarkar	832102,UCIL JADUGUDA STORE CHIEF SUPERINTENDENT STORE AT/PO-JADUGUDA MINES MAIN BUILDING URANIUM CORPORATION OF INDIA LTD Ph No : 9431139916	10	90

Power Contactors As Per IS 13947 - 4 - 1 (10 pieces)

Technical Specifications/तकनीकी विशिष्टियाँ

* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
Dimensions	Length	100.0 (millimeter)
	Width	70.0 (millimeter)
	Height	60.0 (millimeter)
Generic	Auxiliary Contact	1NO
	Mounting	Surface Mounting Via Screw
	AC Rating (Ampere)	16
	Coil Voltage	415
	Contact Voltage	480V AC, 50-60 Hz
	Add on Block	Front
	Duty	AC-3
Warranty Details	Time Duration for Repairing /Replace the defect during Warranty	With Two Week
	Warranty	12.0 (month)
Standards and Certifications	Standard compliance	IEC 60947-4-1, IS 13947-4-1
	Certification	NA
Test Reports	Availability Of Test Report From Central Govt/Nabl/Ilac Accredited Lab To Prove Conformity To Specification	No
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्र	Delivery Days/डिलीवरी के दिन
1	Supriyo Sarkar	832102,UCIL JADUGUDA STORE CHIEF SUPERINTENDENT STORE AT/PO-JADUGUDA MINES MAIN BUILDING URANIUM CORPORATION OF INDIA LTD Ph No : 9431139916	10	90

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**1. Generic**

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Generic

Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 1 year before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

3. Generic

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

- The Seller fails to comply with any material term of the Contract.
- The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
- The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
- The Seller becomes bankrupt or goes into liquidation.
- The Seller makes a general assignment for the benefit of creditors.
- A receiver is appointed for any substantial property owned by the Seller.
- The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

4. Scope of Supply

Scope of supply (Bid price to include all cost components) : Only supply of Goods

5. Past Project Experience

Proof for Past Experience and Project Experience clause: For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed.b. Execution certificate by client with contract value.c. Any other document in support of contract execution like Third Party Inspection release note, etc.**Proof for Past Experience and Project Experience clause:** For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:a. Contract copy along with Invoice(s) with self-certification by the bidder that service/supplies against the invoices have been executed.b. Execution certificate by client with contract value.c. Any other document in support of contract execution like Third Party Inspection release note, etc.

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा

जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो।बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---